

RESEARCH NOTE

Open Access



Carotid artery velocity time integral and corrected flow time measured by a wearable Doppler ultrasound detect stroke volume rise from simulated hemorrhage to transfusion

Jon-Émile S. Kenny^{1,2*} , Igor Barjaktarevic³, David C. Mackenzie^{4,5}, Mai Elfarnawany², Zhen Yang², Andrew M. Eibl^{1,2}, Joseph K. Eibl^{1,2,6}, Chul-Ho Kim⁷ and Bruce D. Johnson⁷

Abstract

Objective: Doppler ultrasonography of the common carotid artery is used to infer stroke volume change and a wearable Doppler ultrasound has been designed to improve this workflow. Previously, in a human model of hemorrhage and resuscitation comprising approximately 50,000 cardiac cycles, we found a strong, linear correlation between changing stroke volume, and measures from the carotid Doppler signal, however, optimal Doppler thresholds for detecting a 10% stroke volume change were not reported. In this *Research Note*, we present these thresholds, their sensitivities, specificities and areas under their receiver operator curves (AUROC).

Results: Augmentation of carotid artery maximum velocity time integral and corrected flowtime by 18% and 4%, respectively, accurately captured 10% stroke volume rise. The sensitivity and specificity for these thresholds were identical at 89% and 100%. These data are similar to previous investigations in healthy volunteers monitored by the wearable ultrasound.

Keywords: Carotid Doppler, Stroke volume, Velocity time integral, Corrected flow time

Introduction

Inferring change in stroke volume (SV) is the bedrock of functional hemodynamic monitoring [1, 2]. Yet measuring SV change (SV_{Δ}) is clinically challenging, so surrogates like the common carotid artery Doppler pulse have been proposed [3, 4] with some conflicting data [5]. We contend that statistically-inadequate beat sampling coupled with variation introduced by the respiratory cycle are important arbiters of inconsistent clinical research

[6]. To rectify the shortcomings of handheld examinations, we developed a wireless, wearable Doppler ultrasound [7, 8]. This device adheres to the neck, maintains a constant insonation angle and accurately measures beat-to-beat changes across multiple cardiorespiratory cycles [6, 7].

In early, proof-of-concept investigations, the carotid artery maximum velocity time integral (VTI) and corrected flow time (ccFT) accurately identified a +10% SV_{Δ} with thresholds of +15% and +2–4%, respectively [9, 10]. More recently, in a human model of hemorrhage and resuscitation comprising approximately 50,000 cardiac cycles, we found a strong, linear correlation between

*Correspondence: jon-emile@heart-lung.org

² Flosonics Medical, 325 Front Street, Toronto, ON M5V 2Y1, Canada

Full list of author information is available at the end of the article





Fig. 1 Picture of wireless, wearable Doppler ultrasound device

SV_{Δ} , and both changing maximum carotid VTI (VTI_{Δ}) and ccFT ($ccFT_{\Delta}$) [8]. While 10% SV_{Δ} associated with 18% VTI_{Δ} and 4.3% $ccFT_{\Delta}$, the *optimal* VTI_{Δ} and $ccFT_{\Delta}$ thresholds for detecting a +10% SV_{Δ} were not investigated. In this *Research Note*, we report these thresholds, their sensitivities, specificities and areas under their receiver operator curves (AUROC).

Main text

Methods

Clinical setting

11 healthy, adult volunteers with no cardiovascular history and who provided written, informed consent were recruited. The study was approved by the Research Ethics Board of the Mayo Clinic (IRB number 19–010,136).

Adherent Doppler system

The U.S. Food and Drug Administration (FDA) cleared, 4 MHz Doppler ultrasound (Fig. 1) (Flosionics Medical, Sudbury, Canada) was placed and the ccFT, maximum and power-weighted, i.e., centroid, VTIs were captured [6, 8–10].

Lower body negative pressure (LBNP) and stroke volume

As previously reported [6], all subjects underwent a 7-stage protocol in duplicate; a non-invasive SV monitor, the Nexfin® (Edwards Lifesciences, Irvine, California), was synchronized with the Doppler monitor for each cardiac cycle.

Statistical analysis

All data was averaged over 10-s windows. SV and carotid Doppler were referenced to resting baseline to model hemorrhage. Cardiac cycles with artifact or during LBNP stage transition were excluded, as described [8]. SV change from the lowest-tolerated LBNP stage back to atmosphere modeled rapid blood transfusion, data from this transition was included.

Each 10-s data point was dichotomized as $\geq +10\% SV_{\Delta}$ or $< +10\% SV_{\Delta}$ based on the non-invasive SV monitor. An equal number of data points were randomly sampled from negative pressure stages without replacement to match the sample size of the positive cases ($\geq +10\% SV_{\Delta}$) for 1000 iterations. Within each iteration, the optimal threshold for each metric was selected using the Youden index. Averages of optimal thresholds and the corresponding sensitivities and specificities were calculated from all iterations and reported as the best final estimations. AUROC was calculated using the subsamples that produced the same optimal thresholds as the final estimation after 1000 iterations.

Results

The median and interquartile ranges for age and BMI were 27 (23–38) years and 23 (21.9–25.6) kg/m^2 , respectively; 39% were female. Vital signs and their change are previously reported [6]. Figure 2 A–D illustrates the progression of mean arterial pressure (MAP), SV, maximum VTI, ccFT and SV for all protocols. The optimal thresholds of % VTI_{Δ} , % $ccFT_{\Delta}$ and their calculation are shown in Fig. 2 E and F. Both +18% VTI_{Δ} and +4% $ccFT_{\Delta}$ were 89% sensitive and 100% specific at detecting $\geq +10\% SV_{\Delta}$. The areas under their receiver operator curves were 0.97 and 0.98, respectively. Though not illustrated, a +20% change in the *centroid* VTI had identical diagnostic characteristics.

(See figure on next page.)

Fig. 2 Hemodynamic data captured during lower body negative pressure (LBNP) and release. Measures from **A–D** are synchronously captured. Each faint line represents a single protocol, while the emboldened line represents the average of all protocols. **A** Stroke volume (SV) percent change during progressively severe LBNP (i.e., hemorrhage model) and release of LBNP (i.e., rapid transfusion model). **B** Mean arterial pressure (MAP) percent change. **C** velocity time integral (VTI) from the wearable Doppler percent change. **D** corrected flow time (ccFT) percent change. **E** The optimal carotid artery maximum VTI threshold for distinguishing $\geq +10\% SV_{\Delta}$. Each data point represents a 10-s average. Prior to subsampling, there were 3596 data points categorized as $< +10\% SV_{\Delta}$ and 598 data points categorized as $\geq +10\% SV_{\Delta}$. The data categorized as $< +10\% SV_{\Delta}$ were randomly subsampled, iteratively 1000 times, to 598 data points (see methods). The sensitivity of maximum VTI is $532/598 = 89\%$ and specificity is $598/598 = 100\%$. **F** The optimal ccFT threshold for distinguishing $+10\% SV_{\Delta}$. Each data point represents a 10-s average. The sensitivity is $532/598 = 89\%$ and specificity is $598/598 = 100\%$




Fig. 2 (See legend on previous page.)

Discussion

In this model of hemorrhage and transfusion, the % VTI_{Δ} and % $ccFT_{\Delta}$ thresholds for detecting a $\geq +10\%$ SV_{Δ} are consistent with earlier proof-of-concept data using a separate paradigm, also in healthy volunteers [9, 10]. With simulated blood transfusion, the 4% $ccFT_{\Delta}$ approximates an absolute $ccFT_{\Delta}$ of 10 ms, comparable to the 7 ms threshold observed in critically-ill patients with undifferentiated shock [4]. The +18% VTI_{Δ} threshold in this investigation is consistent with the slope of the SV_{Δ} - VTI_{Δ} regression line, described previously [8]. Curiously, the +20% threshold for the power-weighted (i.e., centroid) VTI matches the threshold of changing carotid blood flow in septic patients [3]. Assuming static vessel diameter, changing *centroid* velocity best approximates volumetric flow change [8].

Limitations

Our findings have a number of limitations deserving of elaboration. First, from a broad hemodynamic perspective, common carotid blood flow relates to cardiac output (i.e., total blood flow) by the following, general relationship:

$$\text{carotid blood flow} = \text{cardiac output} \times \frac{Z_{\text{whole body}}}{Z_{\text{carotid artery}}}$$

where Z is the impedance to flow for the listed vascular beds. Thus, using the carotid artery as a surrogate for SV_{Δ} implies a constant ratio of whole body-to-downstream carotid impedance. That HR increased and MAP remained constant during the protocol, $Z_{\text{whole body}}$ (e.g., total vascular resistance) likely rose relative to $Z_{\text{carotid artery}}$, especially if downstream internal carotid artery impedance fell by auto-regulation. This could partly explain the relatively large change in carotid VTI (i.e., 18%) observed to detect a 10% SV_{Δ} . Second, though we did not measure carotid diameter to calculate carotid artery flow, the normal carotid artery pressure-diameter relationship is relatively flat above a MAP of 80 mmHg [11, 12]. More specifically, the carotid diameter changes by only 0.2 mm when MAP rises from 80 to 110 mmHg [12]; the average MAP for our subjects during the lowest tolerated LBNP stage was 99 mmHg. Nevertheless, even a 0.2 mm diameter change in a 7 mm carotid artery affects total flow by $\pm 6\%$ [13]. On the other hand, in hypotensive patients, diameter assessment is likely more important; for example, measuring the diameter of the descending aorta improves the sensitivity of Doppler ultrasonographic flow assessment by approximately one-third in the critically-ill [14]. Third, we chose non-invasive pulse contour analysis as a gold standard because it is continuous, user-independent and accurately trends

% SV_{Δ} , which is important for functional hemodynamic monitoring. However, absolute SV measures by non-invasive pulse contour analysis are less adequate, especially in the critically-ill [15, 16]. Still, we believe that the pattern of SV_{Δ} that we observed is valid because it replicates the SV_{Δ} measured in other LBNP investigations [17, 18] and is consistent with the LBNP SV_{Δ} measured by other gold standards including left ventricular outflow tract VTI [19], bioimpedance [20] and bioreactance [21].

In summary, in this large data set of continuously-monitored SV synchronous with carotid Doppler, multiple measures from a wearable ultrasound patch identified +10% SV_{Δ} with high accuracy.

Abbreviations

AUROC: Area under the receiver operator curve; SV: Stroke volume; SV_{Δ} : Stroke volume change; VTI: Velocity time integral; $ccFT_{\Delta}$: Carotid corrected flow time; VTI_{Δ} : Velocity time integral change; $ccFT_{\Delta}$: Carotid corrected flow time change; FDA: Food and Drug Administration; LBNP: Lower body negative pressure; BMI: Body mass index; MAP: Mean arterial pressure; ms: Milliseconds; $Z_{\text{whole body}}$: Whole body impedance; $Z_{\text{carotid artery}}$: Carotid artery impedance; mmHg: Millimeters of mercury.

Acknowledgements

We wish to acknowledge Briana Ziegler, Alex Carlson and Brad Cierzan for their technical assistance, recruitment, scheduling and data management. Additionally, Matt Myers for his assistance in data analysis.

Authors' contributions

JESK: conception, study design, analysis and interpretation, drafting. IB: analysis and interpretation, critical revisions. DCM: analysis and interpretation, critical revisions. ME: data acquisition, analysis and interpretation, critical revisions. ZY: data acquisition, analysis and interpretation, critical revisions. AME: conception, data acquisition, analysis and interpretation, critical revisions. JKE: conception, data acquisition, analysis and interpretation, critical revisions. CHK: data acquisition, study design, analysis and interpretation, critical revisions. BDJ: data acquisition, study design, analysis and interpretation, critical revisions. All authors read and approved the final manuscript.

Funding

This work was supported by the Office of the Assistant Secretary of Defense for Health Affairs, through the Congressionally Directed Medical Research Program under Award No. W81XWH1910591. Opinions, interpretations, conclusions and recommendations are those of the authors and are not necessarily endorsed by the Department of Defense.

Data availability

The datasets used and/or analysed during the current study are available from the corresponding author on reasonable request.

Declarations

Ethics approval and consent to participate

The study was approved by the Research Ethics Board of the Mayo Clinic (IRB number 19-010136); written and informed consent was obtained for all subjects.

Consent for publication

Not applicable.

Competing interests

JESK, ME, ZY, AME, JKE are employees of Flosonics Medical, a start-up building the wearable ultrasound, IB reports consulting fees for GE Healthcare, DCM, CHK and BDJ report no conflicts.

Author details

¹Health Sciences North Research Institute, 56 Walford Rd, Sudbury, ON P3E 2H2, Canada. ²Flosonics Medical, 325 Front Street, Toronto, ON M5V 2Y1, Canada. ³Division of Pulmonary and Critical Care, Department of Medicine, David Geffen School of Medicine at UCLA, Los Angeles, CA, USA. ⁴Department of Emergency Medicine, Maine Medical Center, Portland, ME, USA. ⁵Tufts University School of Medicine, Boston, MA, USA. ⁶Northern Ontario School of Medicine, Sudbury, ON, Canada. ⁷Human Integrative and Environmental Physiology Laboratory, Department of Cardiovascular Diseases, Mayo Clinic, Rochester, MN, USA.

Received: 12 August 2021 Accepted: 21 December 2021

Published online: 10 January 2022

References

- Kenny J-ES, Barjaktarevic I. Letter to the editor: stroke volume is the key measure of fluid responsiveness. *Crit Care*. 2021;25(1):104.
- Kenny J-ÉS. Functional hemodynamic monitoring with a wireless ultrasound patch. *J Cardiothorac Vasc Anesth*. 2021;35(5):1509–15.
- Marik PE, Levitov A, Young A, Andrews L. The use of bioimpedance and carotid Doppler to determine volume responsiveness and blood flow redistribution following passive leg raising in hemodynamically unstable patients. *Chest*. 2013;143(2):364–70.
- Barjaktarevic I, Toppen WE, Hu S, Montoya EA, Ong S, Buhr R, et al. Ultrasound assessment of the change in carotid corrected flow time in fluid responsiveness in undifferentiated shock. *Crit Care Med*. 2018;11:1040–6.
- Beier L, Davis J, Esener D, Grant C, Fields JM. Carotid ultrasound to predict fluid responsiveness: a systematic review. *J Ultrasound Med*. 2020;39(10):1965–76.
- Kenny J-ÉS, Barjaktarevic I, Mackenzie DC, Elfarnawany M, Math ZYB, Eibl AM, et al. Carotid Doppler measurement variability in functional hemodynamic monitoring: an analysis of 17,822 cardiac cycles. *Crit Care Explor*. 2021;3(6):e0439.
- Kenny J-ÉS, Munding CE, Eibl JK, Eibl AM, Long BF, Boyes A, et al. A novel, hands-free ultrasound patch for continuous monitoring of quantitative Doppler in the carotid artery. *Sci Rep*. 2021;11(1):1–11.
- Kenny J-ÉS, Barjaktarevic I, Mackenzie DC, Elfarnawany M, Yang Z, Eibl AM, et al. Carotid Doppler ultrasonography correlates with stroke volume in a human model of hypovolaemia and resuscitation: analysis of 48570 cardiac cycles. *Br J Anaesth*. 2021;127(2):e60–3.
- Kenny J-ÉS, Barjaktarevic I, Mackenzie DC, Eibl AM, Parrotta M, Long BF, et al. Diagnostic characteristics of 11 formulae for calculating corrected flow time as measured by a wearable Doppler patch. *Intensiv Care Med Exp*. 2020;8(1):1–11.
- Kenny J-ÉS, Barjaktarevic I, Eibl AM, Parrotta M, Long BF, Eibl JK, et al. A carotid Doppler patch accurately tracks stroke volume changes during a preload-modifying maneuver in healthy volunteers. *Crit Care Explor*. 2020;2(1):e0072.
- Magder S. The meaning of blood pressure. *Crit Care*. 2018;22(1):257.
- Hansen F, Mangell P, Sonesson B, Länne T. Diameter and compliance in the human common carotid artery—variations with age and sex. *Ultrasound Med Biol*. 1995;21(1):1–9.
- Taylor KJW, Burns PN, Wells PNT. *Clinical Applications of Doppler Ultrasound*, Ch 4. 2nd ed. Raven Press; 1995. p. 94.
- Monnet X, Chemla D, Osman D, Anguel N, Richard C, Pinsky MR, et al. Measuring aortic diameter improves accuracy of esophageal Doppler in assessing fluid responsiveness. *Crit Care Med*. 2007;35(2):477–82.
- Ameloot K, Palmers P-J, Malbrain ML. The accuracy of noninvasive cardiac output and pressure measurements with finger cuff: a concise review. *Curr Opin Crit Care*. 2015;21(3):232–9.
- Saugel B, Hoppe P, Nicklas JY, Kouz K, Körner A, Hempel JC, et al. Continuous noninvasive pulse wave analysis using finger cuff technologies for arterial blood pressure and cardiac output monitoring in perioperative and intensive care medicine: a systematic review and meta-analysis. *Br J Anaesth*. 2020;125(1):25–37.
- Convertino VA, Ludwig DA, Cooke WH. Stroke volume and sympathetic responses to lower-body negative pressure reveal new insight into circulatory shock in humans. *Auton Neurosci*. 2004;111(2):127–34.
- Skytioti M, Søvik S, Elstad M. Respiratory pump maintains cardiac stroke volume during hypovolemia in young, healthy volunteers. *J Appl Physiol*. 2018;124(5):1319–25.
- Vettorello M, Sher S, Santambrogio S, Calini A, Tardini F, Lippi M, et al. Validation of a novel index of hemorrhage using a lower body negative pressure shock model. *Minerva Anesthesiol*. 2016;82(8):839–49.
- Convertino VA, Cooke WH, Holcomb JB. Arterial pulse pressure and its association with reduced stroke volume during progressive central hypovolemia. *J Trauma Acute Care Surg*. 2006;61(3):629–34.
- Bartels SA, Bezemer R, de Vries FJW, Milstein DM, Lima A, Cherpanath TG, et al. Multi-site and multi-depth near-infrared spectroscopy in a model of simulated (central) hypovolemia: lower body negative pressure. *Intensive Care Med*. 2011;37(4):671–7.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Ready to submit your research? Choose BMC and benefit from:

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

At BMC, research is always in progress.

Learn more biomedcentral.com/submissions

